

RESEARCH

Open Access



Evaluation of a haemozoin-based rapid diagnostic test for diagnosis of imported malaria during the phase of prevention of reestablishment in Sri Lanka

Deepika Fernando^{1*}, Priyaleela Thota², Saveen Semege³, Rahuman Booso⁴, David Bell², Kumudunayana T. de A. W. Gunasekera⁵ and Prasad Ranaweera⁵

Abstract

Background: Sri Lanka, an island nation, has eliminated endemic malaria transmission. Maintaining elimination in the continued presence of vectors requires vigilance in screening people travelling from high malaria-risk areas and a rapid response with focal screening for infections identified in the community. Such screening requires accurate and very rapid assays that enable an immediate response. Both microscopy and rapid diagnostic tests (RDTs) have limitations including sensitivity and speed in screening large numbers, while polymerase chain reaction (PCR) is practical only as laboratory confirmation. This study assessed the utility of 'Gazelle', a novel rapid malaria assay based on magneto-optical detection of haemozoin, a by-product of malaria parasite metabolism.

Methods: Between October 2020 and March 2021, two groups of individuals were screened for malaria by four methods, namely, microscopy, Rapid Diagnostic Test (RDT), Gazelle and PCR. Passive case detection was carried out for confirmation of diagnosis amongst individuals suspected of having malaria. Individuals at high-risk of acquiring malaria, namely persons returning from malaria endemic countries, were screened by active case detection.

Results: Of the 440 individuals screened for malaria, nine malaria positives were diagnosed by PCR, microscopy and the HRP2 band of RDT, which included five *Plasmodium falciparum* infections, two *Plasmodium ovale*, and one each of *Plasmodium vivax* and *Plasmodium malariae*. Gazelle correctly detected the *P. vivax*, *P. ovale* and *P. malariae* infections within the 2 min test time, but did not detect two *P. falciparum* infections giving a sensitivity of 77.8%. Specificity was 100%.

Discussion: The Gazelle, a portable bench top device proved useful to screen a large number of blood samples for non-falciparum parasites within 5 minutes of sample input. Species differentiation, and improvement in *P. falciparum* detection, will be important to broaden utility.

Keywords: Malaria, Diagnosis, Rapid diagnostic test, Haemozoin, Microscopy, Polymerase chain reaction

Background

Sri Lanka received malaria-free certification from the World Health Organization (WHO) in September 2016, becoming the second country in the WHO South East Asia region to be declared malaria-free [1]. In the years prior to elimination, frequent epidemics of malaria were

*Correspondence: deepika@parasit.cmb.ac.lk

¹ Department of Parasitology, Faculty of Medicine, University of Colombo, Colombo, Sri Lanka

Full list of author information is available at the end of the article



© The Author(s) 2022. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

recorded, including the massive epidemic of 1934/35. Near elimination state was reached in 1963 when 17 cases of malaria were diagnosed. However, declining financial and political commitment led to resurgence of the disease, which lasted over the next five decades. From 1999, with renewed malaria control policies, the malaria incidence declined until the last case was reported in October 2012 [2]. Prior to elimination, *Plasmodium vivax* infections accounted for over 90% of malaria infections diagnosed. *Plasmodium falciparum* was present but its incidence varied over time and *Plasmodium malariae* was also prevalent till about the late 1960's after which transmission was interrupted. The Anti Malaria Campaign (AMC) began to classify cases as indigenous and imported since 2008 [2]. Following elimination, with the exception of one introduced and one transfusion induced infection which originated from a case of imported malaria reported in 2018 and 2021, respectively [3, 4], there has been no local transmission of malaria in Sri Lanka.

The number of imported malaria cases reported over 5 years, between 2017–2021, included 97 *P. vivax* cases, 84 *P. falciparum*, 26 *Plasmodium ovale*, 6 *P. malariae*, and one mixed infection. The high number of imported malaria cases being diagnosed combined with high receptivity due to the presence of the primary vector *Anopheles culicifacies* and emergence of *Anopheles stephensi* [5] has led to a risk of reestablishment of the disease in the country. This is further complicated by the presence of a non-immune population, declining physician awareness leading to delayed diagnosis of malaria [6] and dwindling financial support to carry out malaria related surveillance activities.

Two malaria case surveillance strategies play a key role to prevent the reestablishment of malaria in Sri Lanka. Passive Case Detection (PCD), which is the detection of malaria cases amongst people who seek health care on their own, usually for fever; and Active Case Detection (ACD), which involves actively searching for malaria infections, such as in people or populations at high risk but who may or may not be obviously ill [7, 8]. Laboratory confirmation of malaria prior to starting anti-malarial treatment is mandatory in Sri Lanka [9].

Microscopy remains the traditional gold standard for malaria diagnosis in Sri Lanka. Microscopy is performed by Public Health Laboratory Technicians (PHLT), who are posted at government hospitals across the country [2]. Challenges exist in maintaining the proficiency of malaria diagnosis via microscopy in the presence of low case numbers [3, 4]. The Carestart™ Pf/Pan Combo Rapid Diagnostic Tests (RDTs) which are purchased from WHO-prequalified suppliers are being used as a supplementary diagnostic method by the AMC. RDTs are

made available at health care institutions, including those where PHLTs are not available, and medical centres at ports of entry where malaria diagnostic services are available 24 h a day. This test has shown good performance for the detection of symptomatic *P. falciparum* (due to high sensitivity of HRP2) and *P. vivax* but a lower sensitivity has been recorded for the pan specific pLDH for *P. falciparum*, *P. ovale* and *P. malariae* [10]. This has to be taken into consideration when using this RDT as a point-of-care test whenever microscopy facilities are not available, especially in instances where HRP2 gene deletion is a possibility irrespective of the parasite density [11]. However, there have been challenges where both tests have been negative in patients who subsequently were found to be positive for malaria in Sri Lanka [4, 12]. Such discrepancies in diagnostic test results are confirmed by polymerase chain reaction (PCR), which allows for a more sensitive and specific detection of malaria parasites. Due to the cost of consumables, training and equipment, the use of PCR has been restricted for confirmation of diagnosis when microscopy and RDT give inconclusive results, or to confirm the species of the parasite. A well-developed laboratory infrastructure is available only at the AMC headquarters and it is not cost effective to use molecular diagnostic methods on a large scale.

Haemozoin, produced when the malaria parasite digests haemoglobin, which is its primary nutrient source found as component of red blood cells, has been identified as a bio-marker for malaria diagnosis given its magnetic and birefringent properties [13–15]. As the parasite denatures haemoglobin, it creates haem, an iron-containing compound that is toxic to the parasite. To overcome this, the parasite converts haem into an insoluble crystalline form called haemozoin. In the Gazelle device, magnets repeatedly align haemozoin perpendicular to a strong magnetic field that is not retained when the field is removed. Haemozoin particles are birefringent and thereby the amount of polarized light that is transmitted through the sample is reduced proportional to the concentration of haemozoin in the sample [13, 16]. Based on this principle, a haemozoin-based, battery operated, malaria diagnostic device (Gazelle) was also evaluated for malaria diagnosis by the AMC.

This study aimed to compare a new haemozoin-detecting device, the Gazelle, with microscopy, Carestart™ RDT and nested PCR for screening and diagnosis of malaria in people suspected of having malaria over a period of six months.

Methods

Study setting and participants

This study was carried out between October 2020 and March 2021 at the Central Laboratory of the AMC

Headquarters, Colombo, Sri Lanka. Two groups of individuals were screened for malaria (a) the first group comprised individuals referred to the AMC Headquarters (AMC HQ) Central Laboratory for confirmation of malaria based on clinical suspicion of disease. The reasons for clinicians suspecting and referring for confirmation of malaria included (i) travel history to a malaria endemic country, (ii) prolonged fever of unknown origin and, (iii) signs and symptoms suggestive of malaria, (b) the second group comprised individuals at high risk of malaria, namely security forces personnel returning from United Nations peacekeeping missions in Africa and individuals repatriated from African destinations who were screened by active case detection (ACD) by the AMC.

Proportion of agreement between the screening test and the gold standard was assumed as 50% to obtain the maximum sample size. Desired precision considered was 5%. α error was taken as 0.05. The calculated minimal sample size was 418 after allowing for a non-response rate of 10% [17].

Sample collection

Two mL of intravenous blood was collected under aseptic conditions by trained nurses in to EDTA tubes after obtaining written informed consent from the participants to be included in the study. Blood samples of individuals suspected of having malaria were referred to the AMC by clinicians in state and government hospitals. In the case of high-risk individuals who had arrived from malaria endemic countries, the samples were collected within 24 h of arrival in Sri Lanka at the port of entry and transported to the AMC Headquarters.

Laboratory diagnosis

The samples were screened for malaria by microscopy, Carestart™ RDT and Gazelle within 24 h of collection and PCR were carried out thereafter for confirmation of diagnosis. The assays were performed in a blinded manner using coded samples.

Microscopy for malaria parasites

Thick and thin blood films were prepared from the collected blood samples. The slides were stained with Giemsa and screened for malaria parasites using microscopy, independently in a blinded manner by a PHLT, who is an expert microscopist and a parasitologist. A smear was interpreted as negative only after examination of 100 fields with an oil immersion lens at 1,000 magnification [18, 19]. If a smear was positive for malaria, at least 100 microscopic fields were screened to confirm the parasite species. Parasite density, expressed as the number of asexual parasites per microlitre of blood was calculate by

dividing the number of asexual parasites for 200 WBCs counted and multiplying it by an assumed WBC density of 8000 cells/ μ L. In cases where there were fewer than 100 asexual parasites per 200 WBCs in smears, quantification was performed against at least 500 WBCs [18, 20]. Parasite densities were calculated by averaging the two counts obtained by the expert microscopist and parasitologist. Blood smears with non-concordant results (differences between the two microscopists in species diagnosis, or differences in parasite density of > 25%) were re-examined by a third, independent expert (WHO Certified Level 1 Microscopist), and parasite densities were calculated by averaging the two most concordant counts.

Rapid Diagnostic Test (Carestart RDT)

CareStart™ Malaria Pf/PAN (HRP2/pLDH) Ag Combo Test, product G0131 (Access Bio Inc. USA) was used for malaria diagnosis. Performance of the tests and interpretation of the results were done according to the manufacturer's instructions and the Standard Operating Procedures prepared by the AMC [21]. Briefly, five microlitre of blood was collected to the specimen transfer device supplied with the RDT kits. This was added to a sample well and then two drops of assay buffer solution was added to the buffer well. The test result was interpreted in 20 min. Individual band reactivity (HRP2 antigen for *P. falciparum* and pan-specific lactate dehydrogenase, pLDH) was recorded for all positive test results.

Magneto-optical detection of malaria parasites by Gazelle

A point-of-care portable magneto-optical device, Gazelle was also used for malaria diagnosis. Gazelle comprises a reader (small, tabletop) and single use disposable cartridges (Fig. 1). Prior to running samples, a positive and negative control was run on the reader to ensure appropriate functioning. As per manufacturer's instructions



Fig. 1 Table top Gazelle reader

[22, 23], 30 μL of the collected blood was taken with a pipette and inserted to the lower chamber of the Gazelle cartridge and thereafter the rest of the sample was stored for molecular diagnosis. 65 μL of the buffer was added to the Gazelle cartridge and the cartridge was inserted into the slot in the reader. The device was run and the results read within a minute and recorded. The individual performing Gazelle was blinded to the results of microscopy and Carestart RDT. Figure 2 depicts the principles of operation of Gazelle.

Nucleic acid amplification assays (nPCR) for diagnosis

To confirm the presence or absence of parasites and parasite species, all samples were analysed by nested PCR assay at the AMC HQ Central Laboratory according to standardized protocols and published methods [24]. In brief, 200 μL of blood was used from whole blood samples. After washing 3 times, DNA was extracted using Qiagen DNA extraction kits according to manufacturer's instructions. Extracted DNA was eluted in 200 μL Buffer AE. The purified DNA (~ 5 μL) was used as a template to detect malaria parasites using genus/species specific primers targeting 18S rRNA gene. Initially a *Plasmodium* genus specific nested PCR was carried out for all samples. Primers rPLU1/rPLU5 were used for the genus specific nest 1 reactions. The product of the first amplification reaction (1 μL) was used as the template for the genus specific second amplification reaction (nest 2) with the oligonucleotide primers rPLU 3 and rPLU 4. The base pair position of these primers with respect to the 18S rRNA A-type gene of *P. falciparum* (GenBank accession no.

M19173) are for rPLU3 132–161 and for rPLU4 353–364 respectively. This gave a PCR product of approximately 235 bp. Whenever the genus-specific nest 2 PCR revealed positive results, the following species-specific nest 2 primers were used to determine the *Plasmodium* species: rFAL1/rFAL2 (*P. falciparum*), rVIV1/rVIV2 (*P. vivax*), rMAL1/rMAL2 (*P. malariae*), rOVA1/rPLU2 (*P. ovale*), and Pmk8/Pmkr9 (*P. knowlesi*). PCR products were electrophoresed for 30 min on a 2% (w/v) agarose gel at constant voltage of 120 V in $0.5 \times$ Tris Borate EDTA buffer. The separated bands were visualized under an ENDuro GDS UV trans illuminator (312 nm wave length) after staining for 15 min with ethidium bromide. Fragment sizes were estimated relative to the 100 bp ladder marker. A sample was considered as positive for *Plasmodium* if a DNA fragment of the expected size of 235 bp was visible. The products of the primary genus specific PCR reaction was used as the template for four separate species specific nested PCR assays.

STARD adherence

STARD (Standards for Reporting of Diagnostic Accuracy Studies) shown in the Additional file 1: Fig. S1 details the study test results, including an analysis of accuracy compared with reference standards for microscopy and PCR (true and false positives and negatives), and comparison of results with RDT as a comparator (concordance and discordance between Gazelle™ and RDT results). The STARD checklist is shown in the Additional file 2: Table S1.




Fig. 2 Principles of operation of Gazelle

Data analysis

The clinical data and laboratory results were collected on record forms and later entered into a Microsoft Excel database. Data recorded on the Gazelle was transferred to the parent company after each batch of samples was run.

The performance of Gazelle was calculated and compared with microscopy, RDT and PCR (the reference standard) with 95% confidence intervals for the following values: sensitivity, specificity, negative likelihood reactions, odds ratio and positive and negative predictive values (PPV, NPV). The total number of samples is those samples on which all four diagnostic tests were successfully performed.

Results

Characteristics of the population

440 individuals were screened for malaria over a six month period between October 2020 and March 2021.

Of the 21 individuals that were screened by passive case detection eight gave a history of travel to a malaria endemic country (i.e. Uganda, Djibouti and Mozambique). The other 13 patients had not visited any overseas country within the past 1 year but were referred to exclude malaria in the differential diagnosis. The balance 419 were classified as a high-risk population for malaria and screened by ACD [7] (Fig. 3). 418 of these individuals were military personnel who had returned after a 12–14 month United Nations peacekeeping mission in African destinations (namely Central African Republic, South Sudan and Mali) and one individual was a returnee from Ethiopia who was at a quarantine centre and screened by ACD.

Malaria detection

Nine malaria positive patients were diagnosed by microscopy (Table 1), which included seven individuals who were screened by PCD and had a travel history




Table 1 Parasitological information of the individuals that were detected by different malaria diagnostic tests

Serial no	Species	Microscopy		Care-Start RDT		Gazelle	PCR
		Density at diagnosis	Stages	HRP2	LDH	Density range	Confirmed species
1	<i>P. vivax</i>	1904/μL	T/G	–	+	Positive	<i>P. vivax</i>
2	<i>P. falciparum</i>	339.8/μL	G	+	+	Positive	<i>P. falciparum</i>
3	<i>P. falciparum</i>	96/μL	R	+	–	Negative	<i>P. falciparum</i>
4	<i>P. falciparum</i>	1188/μL	R	+	–	Negative	<i>P. falciparum</i>
5	<i>P. falciparum</i>	9170/μL	R	+	+	Positive	<i>P. falciparum</i>
6	<i>P. malariae</i>	624/μL	T/S/G	–	+	Positive	<i>P. malariae</i>
7	<i>P. falciparum</i>	54698/μL	R	+	+	Positive	<i>P. falciparum</i>
8	<i>P. ovale</i>	901/μL	T/S/G	–	+	Positive	<i>P. ovale</i>
9	<i>P. ovale</i>	4815/μL	T/S/G	–	+	Positive	<i>P. ovale</i>

R Ring stage, T Trophozoite stage, S Schizont stage, G Gametocyte stage of malaria parasites, HRP2 Histidine Rich Protein 2, LDH Lactate dehydrogenase enzyme

to a malaria endemic country, and two individuals diagnosed during screening by ACD. None of the individuals diagnosed with malaria had taken chemoprophylaxis during their overseas visit.

The nine malaria positive cases included five *P. falciparum*, two *P. ovale* and one each of *P. vivax* and *P. malariae* infections. The results of Carestart RDT (HRP2 band) and PCR were consistent with microscopy (Table 1). Gazelle however gave false negative results in two patients with *P. falciparum* infections.

Comparison of diagnostic test results

The performance of the diagnostics methods tested, microscopy, RDT, PCR and Gazelle were compared. There was no significant difference between the number of samples tested with the positive results (n=9), sensitivity (100%) and specificity (100%) between microscopy, RDT and PCR methods. The results of comparison of Gazelle with other methods are given in Table 2. As compared to microscopy, Gazelle had a sensitivity of 77.8% (100% for non-falciparum species, but missed two out of five *P. falciparum* cases). Specificity was 100% (Table 2). One missed *P. falciparum* case had a parasite density of 96/μl and the other 1188/μl (Table 1). The pLDH band of the Carestart RDT was also negative in these two cases (Table 1).

Table 2 Performance characteristics of diagnostic tests (n = 440)

	Microscopy (gold standard)/ RDT/PCR	
	Positive	Negative
Gazelle		
Positive	7	0
Negative	2	431
Sensitivity (95% CI)	77.8% (40.0% to 97.2%)	
Specificity (95% CI)	100.0% (99.2% to 100.0%)	
PPV(95% CI)	100.0%	
NPV(95% CI)	99.6% (98.5% to 99.9%)	
Accuracy (95% CI)	99.6% (98.4% to 100.0%)	
Odds ratio	NA	
Positive likelihood ratio	NA	
Negative likelihood ratio	0.222	

Positive Likelihood Ratio = Sensitivity/(1-Specificity)

Negative Likelihood Ratio = (1 - Sensitivity)/Specificity

DOR = (TP/FN)/(FP/TN) where DOR is Diagnostic Odds Ratio

The performance of the Gazelle was compared with the other diagnostic methods tested (i.e., microscopy, RDT and PCR). Similar results were obtained for all 3 comparisons. Specificity, sensitivity, PPV, NPV and accuracy (with 95% confidence intervals) for Gazelle as compared to microscopy, RDT and PCR are given in Table 2. Gazelle diagnosed only 7 malaria infections while all other methods detected 9 infections

Discussion

This study aimed to determine the utility of the Gazelle as a rapid screening device for malaria infection, using typical populations where rapid screening would be an advantage. The low positive rate in this population limited the power to determine test accuracy. Compared to microscopy, RDT (HRP2) and PCR, Gazelle did not detect two *P. falciparum* malaria positives. However, during the study period, the Gazelle reader, which is a portable device, weighing less than 5 kg proved rapid and otherwise accurate in the rapid screening of blood samples. The benefits of using Gazelle in this study was identified as the need for low level of user training to perform the test, the ability to screen a large volume of samples in a short period of time and the rapidity of obtaining results.

The failure to detect the two ring stage infections of *P. falciparum* is a concern. Haemozoin has been demonstrated in early ring stages [25], but inability to detect the ring stages by Gazelle could reflect a lower haemozoin concentration in young parasitic stages, as identified elsewhere [26, 27]. Other haemozoin detection systems have claimed detection of *P. falciparum* at parasite densities below 40 parasites/uL at the ring stage [28, 29]. Gazelle’s sensitivity and specificity for diagnosis of low densities of *P. vivax* has also been shown to be similar to microscopy and better than RDTs [30].

Treatment in Sri Lanka is based on species identification and thus, until species identification is possible, Gazelle will not be able to be used as a sole test for diagnosis but it has the potential to be used as a screening test at special locations such as ports of arrival. The developer is developing a species-specific algorithm based on the variation in haemozoin crystal morphology between *Plasmodium* species [31]. Gazelle was accurately sensitive to diagnose both cases of *P. ovale* and *P. malariae* (noting the low numbers of these species in this study) which could prove useful due to the fact that Carestart RDT (pan specific-LDH) shows a lower sensitivity to diagnose both these species in Sri Lanka [10].

The blood samples collected for this study were transported to the AMC central laboratory where Gazelle was carried out. Should the test be carried out at a port of entry, it may prove to be cumbersome due to the requirement of pipettes to measure and insert the required blood volume and diluent into the Gazelle cartridge. This could be overcome by preparation of a one-step cartridge for Gazelle that could directly collect finger prick blood samples. This method is in development and should be available in the market in the near future.

In the absence of its ability to quantify and demonstrate the parasite species, Gazelle will be of advantage to be used in combination with HRP2 Carestart RDTs

as a screening tool to screen large numbers of high-risk individuals arriving at ports of entry. Compared to the RDT which takes approximately 15–20 min, Gazelle can be performed in a very short time hence if the limitation regarding detection of *P. falciparum* can be addressed, it would be a better tool considering the very short time required for testing and giving the result.

Conclusions

Gazelle, a haemozoin based malaria detecting device has potential for rapid throughput malaria screening in Sri Lanka due to its ease of use and availability of results within a short period of time. Sri Lanka is in the phase of prevention of re-establishment, where not only a limited number of malaria cases are diagnosed, but accurate diagnosis of each and every case is essential to prevent onward transmission. Thus, the technical challenges such as species identification, quantifying parasite density and ability to perform the test using finger prick blood should be addressed and overcome if it is to be used as a stand-alone screening device in a country such as Sri Lanka.

Abbreviations

ACD: Active Case Detection; AMC: Anti Malaria Campaign; AMC HQ: Anti Malaria Campaign Headquarters; PCD: Passive Case Detection; PHLT: Public Health Laboratory Technician; POR: Prevention of re-establishment; RDT: Rapid Diagnostic Test; WHO: World Health Organization.

Supplementary Information

The online version contains supplementary material available at <https://doi.org/10.1186/s12936-022-04283-7>.

Additional file 1: Figure S1. STARD Diagram.

Additional file 2: Table S1. STARD checklist for the reporting of studies of diagnostic accuracy. STARD Checklist.

Acknowledgements

The authors wish to acknowledge the following staff of the Anti. Malaria Campaign: Dr. D.P.R.C. Silva, Medical Officer who assisted in coordination of screening, Mr. L. Dissanayake and Mr. M.D.M.S. Marasinghe, Public Health Laboratory Technicians, Mr. H.C.G. Bandara and Mr. Prasanna, Public. Health Field Officers. We also express our gratitude to Mr. M. Thilakarathne, Public. Health Inspector, Sri Lanka Armed Services for all the support provided to collect and transport the blood samples to the Anti Malaria Campaign.

Author contributions

DF, PT and PR planned the study. KG, SS, and RB assisted in data collection. DF and KG supervised the laboratory diagnosis. AR and PT analysed the data. DF prepared the draft manuscript. PT and DB improved on the manuscript. All authors read and approved the final manuscript.

Funding

Hemex Health Inc; USA.

Availability of data and materials

The datasets generated and/or analysed in this publication are not publicly available due to the fact that they belong to the Ministry of Health, Sri Lanka. Clarifications regarding data can be made through Dr. Prasad Ranaweera,

Director of the Anti Malaria Campaign, Sri Lanka who is an author of this publication.

Declarations

Ethics approval and consent to participate

Ethics approval was obtained from the Ethics Review Committee, Faculty of Medicine, University of Colombo (ERC-19–145).

Consent for publication

verbal consent has been obtained from the participants of the study to publish the paper. The individuals screened have not been identified in any manner.

Competing interests

PT, DB and AR are the employees of Hemex Health who funded the study.

Author details

¹Department of Parasitology, Faculty of Medicine, University of Colombo, Colombo, Sri Lanka. ²Hemex Health, Portland, OR, USA. ³Army Headquarters, Sri Jayawardenepura, Sri Lanka. ⁴Directorate of Health Services, Campaign, Sri Lanka. ⁵Anti Malaria Campaign, 555/5 Public Health Building, Narehenpita, Sri Lanka.

Received: 21 April 2022 Accepted: 21 August 2022

Published online: 10 September 2022

References

1. Premaratne R, Ortega L, Janakan N, Mendis KN. Malaria elimination in Sri Lanka: what it would take to reach the goal. *WHO South East Asia J Public Health*. 2014;3:85–9.
2. Premaratne R, Wickremasinghe R, Ranaweera D, de Gunasekera AW, WMKT, Hevawitharana M, Pieris L, et al. Technical and operational underpinnings of malaria elimination from Sri Lanka. *Malar J*. 2019. <https://doi.org/10.1186/s12936-019-2886-8>.
3. Karunasena VM, Marasinghe M, Amarasinghe S, Koo C, Senaratne PAS, Hasantha MBR, et al. The first introduced malaria case reported from Sri Lanka after elimination: implications for preventing the re-introduction of malaria in recently eliminated countries. *Malar J*. 2019. <https://doi.org/10.1186/s12936-019-2843-6>.
4. Chulasiri P, Ranaweera P, Sudarshan P, Jayasinghe M, Harishchandra J, Gunasekera K, et al. Transfusion-induced *Plasmodium falciparum* malaria in a beta thalassaemia patient during the prevention of re-establishment phase in Sri Lanka. *Malar J*. 2021;20:352.
5. Gayan Dharmasiri AG, Perera AY, Harishchandra J, Herath H, Aravindan K, Jayasooriya HTR, et al. First record of *Anopheles stephensi* in Sri Lanka: a potential challenge for prevention of malaria reintroduction. *Malar J*. 2017;16:326.
6. Karunaratna S, Ranaweera D, Vitharana H, Ranaweera P, Mendis K, Fernando D. Thrombocytopenia in malaria: a red-herring for dengue, delaying the diagnosis of imported malaria. *J Global Infect Dis*. 2021;13:172–6.
7. Anti Malaria Campaign, Sri Lanka. National Strategic plan for prevention of reintroduction of malaria in Sri Lanka 2018–2022. http://www.malaria-campaign.gov.lk/images/Publication%20Repository/NSP2018-2022_AMC/National_Strategic_Plan_for_Por___AMC_Sri_Lanka_2018-2022.pdf Accessed 30 Mar 2022.
8. Gunasekera WM, Premaratne R, Fernando D, Munaz M, Piyasena MGY, Perera D, et al. A comparative analysis of the outcome of malaria case surveillance strategies in Sri Lanka in the prevention of re-establishment phase. *Malar J*. 2021;20:80.
9. Anti Malaria Campaign, Sri Lanka. Guidelines on malaria chemotherapy and management of patients with malaria. Office of the Director General of Health Services; 2014. General Circular No. 02–112/2014. http://amc.health.gov.lk/Circulars/Treatment-guidelines_Malaria.pdf Accessed 2nd Jun 2022.
10. de A W Gunasekera WMKT, Premaratne RG, Premawansa WS, Weerasena OVDSJ, Handunetti SM, Fernando SD. Utility of pf/pan RDT for diagnosis

in the prevention of re-establishment of malaria in Sri Lanka. *Pathog Glob Health*. 2018;12:360–7.

11. Cheng Q, Gatton ML, Barnwell J, Chiodini P, McCarthy J, Bell D, et al. *Plasmodium falciparum* parasites lacking histidine-rich protein 2 and 3: a review and recommendations for accurate reporting. *Malaria J*. 2014;13:283.
12. Dharmawardena P, Premaratne RG, Gunasekera WM, Hewavitarane M, Mendis K, Fernando D. Characterization of imported malaria, the largest threat to sustained malaria elimination from Sri Lanka. *Malar J*. 2015;14:177.
13. Mens PF, Matelon RJ, Nour BY, Newman DM, Schallig HD. Laboratory evaluation on the sensitivity and specificity of a novel and rapid detection method for malaria diagnosis based on magneto-optical technology (MOT). *Malar J*. 2010;9:207.
14. Newman DM, Heptinstall J, Matelon RJ, Savage L, Wears ML, Beddow J, et al. A magneto-optic route towards the in-vivo diagnosis of malaria: preliminary results and preclinical trial data. *Biophys J*. 2008;95:994–1000.
15. Nyunt M, Pisciotto J, Feldman AB, Thuma P, Scholl PF, Demirev PA, et al. Detection of *Plasmodium falciparum* in pregnancy by laser desorption mass spectrometry. *Am J Trop Med Hyg*. 2005;74:485–90.
16. Webster M, Kumar V. Gazing at the crystal rods of malaria. *Clin Chem*. 2014;60:1353–5.
17. Lwanga SK, Lemeshow S. Sample size determination in health studies, a practical manual. Geneva: World Health Organization; 1991. p. 1–3.
18. WHO. Basic Malaria Microscopy- part 1: Learners Guide. 2nd Edn. Geneva, World Health Organization, 2010. https://apps.who.int/iris/bitstream/handle/10665/44208/9789241547826_eng.pdf;jsessionid=5375D9DA0A7CC7AF412A102B30D3F4CE?sequence=1. Accessed 30 Mar 2022
19. Anti Malaria Campaign, Sri Lanka. Manual for Parasitological Surveillance in prevention of reintroduction/reestablishment of malaria in Sri Lanka. 2019. <http://www.malariacampaign.gov.lk/images/PublicNotice-Repository/Manual-for-Parasitological-Surveillance-in-prevention-of-reintroduction--or-reestablishment-of-malaria-in-Sri-Lanka.pdf> Accessed 30 Mar 2022.
20. WHO. Malaria parasite counting. Malaria microscopy standard operating procedure. MM-SOP-09. Version 1. World Health Organization, Geneva, 2016. <https://www.who.int/publications/i/item/HTM-GMP-MM-SOP-09> Accessed 30 Mar 2022.
21. Anti Malaria Campaign, Sri Lanka. Standard operating procedure for rapid diagnostic test kits. 2015. <http://www.malariacampaign.gov.lk/images/Publication%20Repository/SOP/SOP%20RDT.pdf> Accessed 30 Mar 2022.
22. de Melo GC, Netto RLA, Mwangi VI, Salazar YEAR, de Sampaio S V, Monteiro WM, et al. Performance of a sensitive haemozoin-based malaria diagnostic test validated for vivax malaria diagnosis in Brazilian Amazon. *Malar J*. 2021;20:146.
23. Rajat Kumar AKV, Shrivastava S, Thota P, Singh MP, Rajasubramaniam S, Das A, Bharti PK. First successful field evaluation of new, one-minute hemozoin-1 based malaria diagnostic device. *EclinMed*. 2020;22: 100347.
24. Snounou G, Singh B. Nested PCR analysis of *Plasmodium* parasites. *Methods Mol Med*. 2002;72:189–203.
25. Molnár P, Orbán Á, Izrael R, Babai R, Marton L, Butykai Á, et al. Rapid and quantitative antimalarial drug efficacy testing via the magneto-optical detection of hemozoin. *Sci Rep*. 2020;10:14025.
26. Rebelo M, Shapiro HM, Amaral T, Melo-Cristino J, Hänscheid T. Haemozoin detection in infected erythrocytes for *Plasmodium falciparum* malaria diagnosis—prospects and limitations. *Acta Trop*. 2012;123:58–61.
27. Delahunt C, Horning MP, Wilson BK, Proctor JL, Hegg MC. Limitations of haemozoin-based diagnosis of *Plasmodium falciparum* using dark-field microscopy. *Malar J*. 2014;13:147.
28. Orban A, Butykai A, Molnar A, Prohle Z, Fulop G, Zelles T, et al. Evaluation of a novel magneto-optical method for the detection of malaria parasites. *PLoS ONE*. 2014;9: e96981.
29. Baptista V, Peng WK, Minas G, Veiga MI, Catarino SO. Review of micro-devices for hemozoin-based malaria detection. *Biosensors*. 2022;12:110.
30. Valdivial HO, Thota P, Braga G, Ricopa L, Barazorda K, Salas C, et al. Field validation of a magneto-optical detection device (Gazelle) for portable point-of-care *Plasmodium vivax* diagnosis. *PLoS ONE*. 2021;16: e0253232.
31. Noland GS, Briones N, Sullivan DJ Jr. The shape and size of hemozoin crystals distinguishes diverse *Plasmodium* species. *Mol Biochem Parasitol*. 2003;130:91–9.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

